
p-DICHLOROBENZENE



1. PERFORMANCE

1) Measuring range $\begin{array}{c} \hbox{:} \ 0.01\text{-}0.40 \ ppm \\ \hbox{:} \ 200 \text{m}\ell/\text{min} \times 15 \text{min} \\ \end{array}$

3) Detectable limit : 0.002 ppm4) Shelf life : 1 year5) Operating temperature $: 10 \sim 35 \text{ }^{\circ}\text{C}$

6) Temperature compensation : Necessary (See "TEMPERATURE CORRECTION TABLE")

7) Reading : Direct reading from the scale calibrated at the sampling of $200 \text{m} \ell \times 15 \text{min}$

8) Colour change : Orange → Reddish purple

2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 10% RSD-high: 10%

3. CHEMICAL REACTION

By reacting with an Oxidizer, Hydrogen chloride is produced and PH indicator is discoloured. $C_6H_4Cl_2 + PbO_2 + H_2SO_4 \rightarrow HCI$

4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence			
Trichloroethylene	Similar stain is produced.	Higher readings are given.			
Tetrachloroethylene	"	"			
1,2-Dichloroethylene	"	"			
Vinyl chloride	"	"			

(NOTE)

Air sampler is required for this tube.

TABLE OF THE COEFFICIENT FOR TEMPERATURE CORRECTION(20 °C standard)

	Temp(°C)	0	1	2	3	4	5	6	7	8	9
	10	2.13	1.95	1.78	1.63	1.50	1.38	1.28	1.19	1.11	1.05
	20	1.00	0.95	0.92	0.88	0.84	0.81	0.78	0.75	0.73	0.72
	30	0.70	0.69	0.68	0.68	0.67	0.66	_	_	_	_